

**COOLER HEATER EQUIPMENT RFP****FORMAL REQUEST FOR PROPOSAL****ADDENDUM NUMBER: ONE (1)****MARCH 19, 2021**

**THIS ADDENDUM IS ISSUED PRIOR TO THE ACCEPTANCE OF THE FORMAL RFPs. THE FOLLOWING CLARIFICATIONS, AMENDMENTS, ADDITIONS, DELETIONS, REVISIONS, AND MODIFICATION FORM A PART OF THE CONTRACT DOCUMENTS ONLY IN THE MANNER AND TO THE EXTENT STATED.**

**Q&A**

1. VENDOR XYZ does not offer a dual pump device. As a result, to get the functionality your perfusion team would require, we will be bidding (6) devices, not (3). Is that an acceptable offer for your consideration? **No this is not acceptable; Broward Health cannot support two units in the OR. The equipment must be able to heat/cool patient and heat /cool cardioplegia within the same unit.**

**END ADDENDUM ONE**